

OCT 04 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130124 ✓
Name		NICHOLAS ULINEMI	ID# 3429 ✓ Date 9/30/10 ✓
A	Agency	METLAKATLA POLICE DEPT.	Phone # 886-4011
Instrument Location		METLAKATLA POLICE DEPT.	
Alco S/N		X173012 ✓	Target Value .086 ✓ High Pressure .086
B	Alco Test Values	.088 ✓ 1 st Alco	.088 ✓ 2 nd Alco
Signature		Nicholas Ulinemi	
		(OVER)	

BMB
10/5/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130124

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

SEPTEMBER 30, 2010

OPERATOR'S NAME:

VLINIEH/NICHOLAS/L

OPERATOR'S NUMBER: 3429

SUBJECT'S LAST NAME:

VER

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 0000

DEPT/AGENCY: MET1

CASE/REPORT: 09999

TEST TYPE: U

ALCO TARGET VALUE: .006

ALCO S/N: X173012

— BREATH ANALYSIS —

.006 ADJUSTED FOR 30.16 in		
ALCO TARGET	.006	13:09
BLANK TEST	.000	13:10
INTERNAL STANDARD	VERIFIED	13:10
ALCO TV 30.17 in	.008	13:11
BLANK TEST	.000	13:12
SUBJECT SAMPLE	.000	13:12
BLANK TEST	.000	13:13
ALCO TV 30.16 in	.008	13:13
BLANK TEST	.000	13:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

SEPTEMBER 30, 2010

TIME 13:07

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	30.16 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () * + , - .